

## Amount of substance, pH, Hong Kong (China), GL (Government Laboratory )



NMI Service Identifier	Measurement Service Category	Matrix	Quantity	Dissemination Range of Measurement Capability		Range of Expanded Uncertainties as Disseminated			Mechanism(s) for Measurement Service Delivery	Comments
				From	To	From	To	Is the expanded uncertainty a relative one?		
GLHK017	pH	buffer solution	pH	6.86	6.89	0.011	0.011	No	Calibration service	Approved on 06 december 2011

The expanded uncertainties correspond to  $k = 2$  (level of confidence 95%)