

Electricity and Magnetism, Latvia, LATMB (Latvian Metrology Bureau)



Calibration or Measurement Services			Measurand Level or Range			Measurement Conditions/Independent variables		Expanded Uncertainty						Comments	NMI Service Identifier
Quantity	Instrument or artifact	Instrument Type or Method	Minimum value	Maximum value	units	Parameter	Specifications	Value	Units	Coverage Factor	Level of Confidence	Is the expanded uncertainty a relative one?	Uncertainty matrix		
DC voltage sources: single values	Solid state voltage standard	Direct comparison	10	10	V			4.9	μV	2	95%	No		Approved on 14 March 2006	1
DC voltage sources: single values	Standard cell, solid state voltage standard	Direct comparison	1.018	1.018	V			5.1	μV	2	95%	No		Approved on 14 March 2006	2
DC resistance standards and sources: low values	Fixed resistor	DCC bridge	0.001	0.001	Ω			0.0042	$\mu\Omega$	2	95%	No		Approved on 14 March 2006	13
DC resistance standards and sources: low values	Fixed resistor	DCC bridge	0.01	0.01	Ω			0.024	$\mu\Omega$	2	95%	No		Approved on 14 March 2006	14
DC resistance standards and sources: low values	Fixed resistor	DCC bridge	0.1	0.1	Ω			0.13	$\mu\Omega$	2	95%	No		Approved on 14 March 2006	15
DC resistance standards and sources: low values	Fixed resistor	DCC bridge	1	1	Ω			1.3	$\mu\Omega$	2	95%	No		Approved on 14 March 2006	16
DC resistance standards and sources: intermediate values	Fixed resistor	DCC bridge	10	10	Ω			13	$\mu\Omega$	2	95%	No		Approved on 14 March 2006	17
DC resistance standards and sources: intermediate values	Fixed resistor	DCC bridge	100	100	Ω			0.13	$\text{m}\Omega$	2	95%	No		Approved on 14 March 2006	18
DC resistance standards and sources: intermediate values	Fixed resistor	DCC bridge	1000	1000	Ω			1.3	$\text{m}\Omega$	2	95%	No		Approved on 14 March 2006	19
DC resistance standards and sources: intermediate values	Fixed resistor	DCC bridge	10	10	$\text{k}\Omega$			22	$\text{m}\Omega$	2	95%	No		Approved on 14 March 2006	20

Electricity and Magnetism, Latvia, LATMB (Latvian Metrology Bureau)



Calibration or Measurement Services			Measurand Level or Range			Measurement Conditions/Independent variables		Expanded Uncertainty						Comments	NMI Service Identifier
Quantity	Instrument or artifact	Instrument Type or Method	Minimum value	Maximum value	units	Parameter	Specifications	Value	Units	Coverage Factor	Level of Confidence	Is the expanded uncertainty a relative one?	Uncertainty matrix		
DC resistance standards and sources: intermediate values	Fixed resistor	DCC bridge	100	100	k Ω			0.23	Ω	2	95%	No		Approved on 14 March 2006	21
DC resistance standards and sources: intermediate values	Fixed resistor	Multimeter	1	1	M Ω			44	Ω	2	95%	No		Approved on 14 March 2006	22
DC resistance standards and sources: high values	Fixed resistor	Multimeter	10	10	M Ω			470	Ω	2	95%	No		Approved on 14 March 2006	23
DC resistance standards and sources: high values	Fixed resistor	Multimeter	100	100	M Ω			45.2	k Ω	2	95%	No		Approved on 14 March 2006	24
AC voltage up to 1000 V: meters	AC voltmeter, multimeter, multifunction transfer standard	Direct with calibrator	0.03	700	V	Frequency	45 Hz to 500 kHz	0.03 to 4.9	%	2	95%	Yes	Matrix5.2.2	Approved on 14 March 2006	45
AC current up to 100 A: meters	AC ammeter, multimeter, multifunction transfer standard	Direct with calibrator	0.029E-03	11	A	Frequency	45 Hz to 10 kHz	0.1 to 1.7	%	2	95%	Yes	Matrix6.2.2	Approved on 14 March 2006	46

Electricity and Magnetism, Latvia, LATMB (Latvian Metrology Bureau)



Uncertainties table: Matrix5.2.2

AC voltage up to 1000 V: meters, LATMB Internal Identifier: 45

	45 Hz	1 kHz	5 kHz	8 kHz	10 kHz	18 kHz	20 kHz	50 kHz	90 kHz	100 kHz	450 kHz	500 kHz
30 mV	0.71	0.71	-	-	0.91	-	1.43	1.69	-	4.9	4.9	-
100 mV	0.05	0.05	-	-	0.07	-	0.09	0.3	-	0.68	0.68	0.68
300 mV	0.24	0.24	-	-	0.44	-	0.68	1.06	-	3.13	3.13	3.13
1 V	0.03	0.03	-	-	0.04	-	0.1	0.3	-	0.3	0.6	-
3 V	0.17	0.17	-	-	0.33	-	0.6	1.07	-	1.07	2.47	-
10 V	0.03	0.03	-	-	0.07	-	0.13	0.3	0.3	-	-	-
30 V	0.2	0.2	-	-	0.38	-	0.79	1.1	1.1	-	-	-
100 V	0.04	0.05	-	-	0.08	0.08	-	-	-	-	-	-
300 V	0.27	0.37	-	-	0.45	0.45	-	-	-	-	-	-
700 V	0.07	0.25	0.41	0.41	-	-	-	-	-	-	-	-

The expanded uncertainties given in this table are expressed in %.

Electricity and Magnetism, Latvia, LATMB (Latvian Metrology Bureau)



Uncertainties table: Matrix6.2.2

AC current up to 100 A: meters, LATMB Internal Identifier: 46

	45 Hz to 65 Hz	45 Hz to 1 kHz	65 Hz to 500 Hz	500 Hz to 1 kHz	1 kHz to 5 kHz	5 kHz to 10 kHz
0.029 mA to 0.330 mA	-	0.48	-	-	0.66	1.7
0.330 mA to 3.3 mA	-	0.19	-	-	0.34	0.84
3.3 mA to 33 mA	-	0.17	-	-	0.31	0.84
33 mA to 330 mA	-	0.17	-	-	0.31	0.84
0.330 A to 2.2 A	-	0.16	-	-	0.90	-
2.2 A to 11 A	0.10	-	0.15	0.41	-	-

The expanded uncertainties given in this table are expressed in %.