

Amount of substance, Inorganic solutions, Argentina, INTI (Instituto Nacional de Tecnologia Industrial)

The expanded uncertainty range spans from the smallest numerical value of the uncertainty to the largest numerical value of the uncertainty found within the quantity range.

NMI Service Identifier	Measurement Service Sub-Category	Matrix	Measurand		Dissemination Range of Measurement Capability			Range of Expanded Uncertainties as Disseminated						Mechanism(s) for Measurement Service Delivery	Comments
			Analyte or Component	Quantity	From	To	Unit	From	To	Unit	Coverage factor	Level of confidence	Is the expanded uncertainty a relative one?		
INORGIS-1	Elemental	acid solution	zinc	Mass concentration	0.5	1000	mg/L	2	10	%	2	95%	Yes	Analytical service and PT	Approved on 27 March 2013
INORGIS-2	Elemental	acid solution	mercury	Mass concentration	0.01	2	mg/L	10	15	%	2	95%	Yes	Analytical service and PT	Approved on 27 March 2013
INORGIS-3	Elemental	acid solution	chromium	Mass concentration	0.5	1000	mg/L	5	25	%	2	95%	Yes	Analytical service	Approved on 27 March 2013
INORGIS-4	Anionic	calibration solution	nitrite	Mass fraction	100	10000	mg/kg	0.5	1.0	%	2	95%	Yes	Analytical service	Approved on 27 March 2013
INORGIS-5	Anionic	calibration solution	nitrate	Mass fraction	100	10000	mg/kg	0.55	1.0	%	2	95%	Yes	Analytical service	Approved on 27 March 2013