

## Amount of Substance, Gases, Serbia, DMDM (Directorate of Measures and Precious Metals)



NMI Service Identifier	Measurement Service Sub-Category	Matrix	Measurand		Dissemination Range of Measurement Capability			Range of Expanded Uncertainties as Disseminated						Mechanism(s) for Measurement Service Delivery	Comments
			Analyte or component	Quantity	From	To	Unit	From	To	Unit	Coverage factor	Level of confidence	Is the expanded uncertainty a relative one?		
GMH-RU-H09	Environmental	purified air	ozone	Amount-of-substance fraction	0	1000	nmol/mol	Q[1.1, 0.022 x(O <sub>3</sub> )]	nmol/mol	2	95%	No	Instrumental comparison against primary UV photometer Calibration	Approved on 28 June 2016	